

Product Return Form - Calibration



Company:	
Name:	
Tel:	
Fax:	
Address:	
e-mail:	

Gemini Data Loggers(UK) Ltd,
Scientific House, Terminus Road,
Chichester, PO19 8UJ, UK.

t +44 1243 813000
f +44 1243 531948
e sales@gemindataloggers.com
w www.gemindataloggers.com

Customer Reference/Order Number:	
---	--

Number of Returned Items:	
----------------------------------	--

Unit Serial Number	Logger Type/Part Number (e.g. TG-0050)	Calibration Points (tick as appropriate)		Standard calibration points are 0 and 30 Degrees Centigrade and 20 and 80% Relative Humidity respectively. These points have been chosen to give good coverage for general purpose use. Should you wish to specify other points see note 3. and specify your chosen points in this column
		Standard	Other	

PLEASE NOTE:

1. If the calibration is for a logger with a probe, this must be returned with the logger. Without the probe Gemini will be unable to calibrate the unit until the probe is supplied.
2. All goods must be returned clean and free from potentially hazardous contamination. Gemini reserves the right not to service or repair goods which do not meet this condition.
3. Calibration certificates can be issued for temperatures between -20°C and +150°C and for humidity between 20% and 90% RH at 25°C.
4. Standard Terms and Conditions apply. Refer to www.gemindataloggers.co.uk for details.